

LEVELS OF MANUAL WHEELCHAIRS

Standard Wheelchair - (K0001 & K0002)

Usually short-term rentals.

Features

- Back Height 16" (fixed)
- Seat Width 16" - 18" - 20"
- Seat Depth 16"
- Seat to Floor 17.5" - 19.5" (K0001 = 19" K0002 = 18" & under)
- Weight Capacity 250#



Lightweight Wheelchair - K0003 (Under 34 lbs.)

Provided for those with diminished upper body strength.

Features

- Back Height 16"
- Seat Width 14" - 22"
- Seat Depth 16"
- Seat to Floor 17.5" - 19.5"
- Weight Capacity 250# (300# on Seat Width 20" & 22")



High-Strength Lightweight Wheelchairs - K0004 (Under 32 lbs.)

Provided for those with diminished upper body strength who need a chair for:

1. Long periods of use to accomplish mobility-related activities of daily living (MRADL), **OR**
2. Need features unavailable on standard or lightweight wheelchair, **OR**
3. Cannot propel any standard or lightweight wheelchair.

Features

- Back Height 15" - 19" (Adjustable)
- Seat Width 14" - 22"
- Seat Depth 16" - 17" - 18"
- Seat to Floor 13.5" - 19.5"
- Weight Capacity 250#



Ultra Lightweight Wheelchairs - K0005 (Under 30 lbs.)

Provided for very active individuals who utilize their wheelchair throughout the entire day.

Features

- Variety of frame types & sizes (12" to 20" width)
- Lightest weight manual wheelchair
- Wheels - Quick Release Axles
- Fixed or swing-away front rigging
- Weight Capacity 250#



Manual Tilt-n-Space Wheelchairs - E1161

Caregiver assistance required to propel and tilt wheelchair. This wheelchair can be configured to meet a variety of seating and positionin needs. Provided for patients who:

1. Cannot perform independent weight shift, **OR**
2. Cannot sit upright in a standard or reclining wheelchair, **OR**
3. Problem swallowing, **OR**
4. Are at risk or currently experiencing skin breakdown.

Features

- Seat to Floor 14" - 19.25"
- Weight Capacity 250# (Heavy-Duty Option 350# max.)



BLACKBURN'S

Our Locations

Corporate Headquarters Showroom/Pharmacy

301 Corbet Street
Tarentum, PA 15084
800-472-2440
724-224-9124 (fax)

Erie Showroom

308 East Sixth Street
Erie, PA 16507
800-328-9325
814-454-2706 (fax)

Harrisburg, PA Regional Office

6345 Flank Drive, Ste 1800
Harrisburg, PA 17112
717-996-3390
717-695-4282 (fax)

Rochester, NY Regional Office

400 Air Park Drive, Ste 100
Rochester, NY 14624
877-472-0192
716-688-9193 (fax)



Serving Regions of Pennsylvania, Ohio, New York and West Virginia

www.blackburnsmed.com

REV: BB_Rehab_082025

MANUAL WHEELCHAIRS ACCESSORIES

The following list is a sampling of the most common attachments and accessories utilized on a manual wheelchair. The patient's medical condition evaluated by the physician and a clinically trained rehab mobility specialist determine the justification for the manual wheelchair configuration.

Cushions and Backs basic, pressure relieving, positioning or custom molded.

Height-Adjustable Arm Rest it is indicated for coverage when the following two criteria are met; Patient requires an arm height different than that available using non-height adjustable armrests, **AND** Patient spends at least two hours per day in the wheelchair.

Rear Anti-Tippers safety feature to prevent rearward tipping of wheelchair.

Safety/Seat Belt patient has weak upper body strength, upper body instability or muscle spasticity which requires use of this item for proper positioning.

Adjustable Angle Footrest patient with high tone, ankle dorsi or pantar flexion, etc. Required to properly position feet on footplate to avoid risk of falling off or causing injury while using wheelchair.

Heel Loops required to keep patient's feet on footplates.

Wheel Lock Extensions assist those who suffer from diminished strength due to stroke, etc. Allows patient to lock wheels independently.

Elevating Leg Rests relieve swelling due to edema, or fracture, or has a reclining back in wheelchair.

Stump Support provides support for leg amputee.

Arm Troughs it is indicated for coverage when one of the following is met;
Patient has quadriplegia, **OR** patient has paraplegia, **OR** has uncontrollable arm movements

Headrest with Adjustable/Removable Hardware necessary above the fixed hardware so it can be adjusted to the correct position and removed for safe transferring or transporting in a vehicle.

Oxygen Holder mobile carrier for oxygen tanks.

One Arm Drive Attachment required to propel wheelchair when only one arm is functional.

Attendant Wheel Lock required above and beyond standard wheel locks because patient is unable to access each wheel independently on manual tilt wheelchair, and attendant is responsible to lock the wheelchair in place for safe transfer.

Solid Tires patient who cannot maintain pneumatic tires.

Solid Seat Pan sits on top of sling seat upholstery to provide a solid sitting surface for use with any type of wheelchair cushion.



RESNA

Rehabilitation Engineering and Assistive
Technology Society of North America

 **BLACKBURN'S**

800-472-2440

Serving Regions of Pennsylvania, Ohio, New York and West Virginia

WWW.**blackburnsmed**.com